


Calibration Data



DROPLET
MEASUREMENT
TECHNOLOGIES

5710 Flatiron Parkway Suite B, Boulder CO 80301, ph. 303-440-5576, www.dropletmeasurment.com

May 2009 WMT PCASP-200 S/N 0102



**DROPLET
MEASUREMENT
TECHNOLOGIES**

5710 Flatiron Parkway Suite B
Boulder, CO 80301
www.dropletmeasurement.com

Calibration Certificate

Customer: WMI	Instrument Model: PCASP 100X	
Serial Number: 0804-0102	Calibration Date: 5/6/2009	

Condition Received

In Tolerance:		Operational Failure:	
Out of Tolerance:	X	Physical Damage:	
New:		Comments:	

Condition Shipped

Meets DMT Specifications:	X	Other:
Calibration Done By:	James Wiggins	
Recommended Re-Calibration Date:	5/6/2010	

NIST-Traceable Particles


Size	Tolerance	Lot No.	Material	Manufacturer	Exp. Date
.12um	<=12%CV	28964	PSL	Duke	10/08
.144um	<=12%CV	28625	PSL	Duke	7/08
.222um	<=3%CV	33078	PSL	Duke	3/11
.43um	<=3%CV	34155	PSL	Duke	10/11
.82um	<=3%CV	33176	PSL	Duke	3/11
2.001um	+.025um	25785	PSL	Duke	3/05

Technician Signature:

Date: 5/6/2009

DROPLET MEASUREMENT TECHNOLOGIES, INC.

5710 Flatiron Parkway, Suite B
Boulder, Colorado 80301
(303) 440-5576

**PCASP Calibration Report**

DMT Order Number:	0901096	Client/Probe Owner:	WMI
Date:	4/8/2009	Model Number:	PCASP 100X
Work Performed by:	James Wiggins	Serial Number:	0804-0102
Other information:None			

PRE-CALIBRATION BEFORE CLEANING

Preliminary Visual Inspection	
General Condition of the Exterior:	Ok
General Condition of the Interior:	Potentiometers on Analog Mux and Baseline Restoration Module had been adjusted after calibration.
Other information:	

Working Status Of Probe Prior To Cleaning		
Condition and general appearance of the laser firing:	Good	
Vibrational sensitivity of the laser:	OK	
Heater Status:	Power Rating (watts):	215
	Voltage Rating (volts):	28
	Calculated Heater Resistance R = V^2/P (ohms)	3.65
	Actual Measured Heater Resistance (ohms)	3.8
+5 V Power Supply:	5.02 VDC	200 mVAC noise
+15 V Power Supply:	15.2 VDC	152 mVAC noise
-15 V Power Supply:	-15.2 VDC	120 mVAC noise

INITIAL CALIBRATION PROCEDURES

A precision bead pre-calibration, using the DMT aerosol generator, was performed on the probe, the results of this pre-calibration are attached to this report.

Our technician's comments about this calibration are as follows:

No particles were peaking and the ambient distribution looked terrible. I checked the laser output coupler and it was filthy but the instrument was still reading above 9 VDC on reference. I cleaned the output coupler and the laser reference immediately went to 10.5VDC which is the maximum value for this housekeeping parameter. I found that someone had adjusted the potentiometer on the multiplexer module to make it look like the probe was okay. After cleaning the particles started to bin correctly and ambient distribution started looking normal.

The sample flow and the sheath flow were measured using pneumatic flow meters and the results are tabulated below. The probe was adjusted for the proper flow rate and the rates were measured and tabulated again.

Time of Measurement	Sample Flow in cc/sec	Sheath Flow in cc/sec
Before Adjustments	1.04	12.54
After Adjustments	1.26	15.2

Additional comments made by the technician performing this procedure are:

The reference voltage was measured to be (See notes above) volts, in a properly operating probe, this voltage should be greater (in magnitude) than six (6) volts.

The signals GAIN1, GAIN3 and GAIN4 were measured and then adjusted to be within the manufacturer's specification. The results of these measurements are tabulated below.

Signal Name:	GAIN1 (mvolts) (Signal-3, Low Gain)			GAIN3 (mvolts) (Signal-2, Mid Gain)			GAIN4 (mvolts) (Signal-1, High Gain)		
	Low Peak	High Peak	Average	Low Peak	High Peak	Average	Low Peak	High Peak	Average
Before Adjustment	274	294	284	232	266	249	-58	68	.05
After Adjustment	328	376	352	320	374	347	-44	328	142

Our technician's comments concerning this procedure are:

CLEANING

A thorough cleaning was performed on the probe, the following table summarizes the steps that were performed.

Electrical Contacts	
Contact Description	Status and/or Service Performed
Amphenol Connector	Checked
Card Seating	Cleaned
Card Edge Connectors	Cleaned
Optical System Components	
Component Description	Status and/or Service Performed
Aspheric Collector	Cleaned
45 Degree Mirror	Cleaned
Parabolic Mirror	Cleaned
Laser Output	Cleaned

Additional comments made by the technician performing this procedure are:

ALIGNMENT

The reference voltage was measured and various adjustments were performed to make it come into specification. Laser Reference Voltage: 9.1

Using the DMT aerosol generator, the probe's inlet jets were adjusted to provide the strongest signal possible to the incoming aerosols.

The desiccant was replaced with fresh royal blue desiccant.

Our technician's comments concerning this procedure are:

CALIBRATION

A final calibration, using the DMT aerosol generator, was performed on the probe, the results of this final calibration are attached to this report.

Our technician's comments about this calibration are as follows:

During final calibration the transformer / voltage regulator combo that drove the laser died. Since it was 24 years old I doubt the laser caused this failure. This coupled with the fact that the probe has ran for several days with the new driver supports that assumption.

Additional comments made by our technician:

I was in contact with the end user a few months ago when he initially had trouble with the probe. He and I spoke and emailed each other to try and troubleshoot the probe in the field. I suggest that the next time the instrument returns from the field that all potentiometers are inspected for tampering. There are four that need to be checked. The three on the baseline restoration module, one on the analog multiplexer, and one on the high voltage power supply near the pump relay.

DHT Particle Analysis and Display System

Program Configure

Sample Read a File

Current Data File: 20090506114322

(0) SPP_200 Setup Debug

Pump Relay Control This light indicates that PADS thinks it has turned the relay on.

Relay On ● There is no direct feed-back from the probe.

SPP_200 Data

Number Conc (# / cm ³)	Sample Flow (vol cm ³ /s)	SPP_200 # Conc (#/cm ³)
2250.25	1.26	
Volume Conc (um ³ / cm ³)	Sheath Flow (vol cm ³ /s)	
7.17	14.19	
MVD (um)	Laser Ref (V)	
0.24	9.04	
ED (um)	Electronics	
0.21	Temp (C)	
Aux Analog 1	Avg Transit	
0	228	
Hi Gain Baseline (V)	ADC Overflow	
0.1	0	
Mid Gain Baseline (V)		
0.34		
Lo Gain Baseline (V)	Pump Flow OK	
0.34		

0cc

00 d 11:46:12 (42372sec)

Slow

Medium

Fast

11:43:23

12:30:00

12:57:50


No Fault v2.7.2

Enabled

Selectable Charts


Read a File

SPP_200 volume Conc (um³/cm³)



COM Port 5

Auto-Scale



Bins

Particle Size

Y-Axis

Log-Scale

Normalized

X-Axis

Auto-Scale

5. DMT Particle Analysis and Display System

Program Configure

Sample Read a File

Current Data File: 20090506114322

(0) SPP_200 Setup Debug

Pump Relay Control This light indicates that Conc/Vol Conc

ON OFF PADS thinks it has turned the relay on.
SPP_200 Data back from the probe.

Number Conc Sample Flow
(# / cm³) (# vol cm³/s)

1809.69 1.26
Volume Conc Sheath Flow
(um³ / cm³)(vol cm³/s)

7.61 14.15
MVD (um) Laser Ref (V)

0.25 9.08
ED (um) Electronics

0.23 39.1
Aux Analog 1 Temp (C)

-0.01 251
Hi Gain

Baseline (V) ADC Overflow
0.07 0

0.34 Pump Flow
Mid Gain OK
Baseline (V)
0.34



0.34 Pump Flow
Lo Gain OK
Baseline (V)
0.34

00 d 11:45:54 (42355sec)

12:30:00 12:57:50

Sample Read a File
Manual Slow Medium Fast
(0) SPP_200 Setup Debug

No Fault v2.7.2



D.U.T Particle Analysis and Display System

Program Configure

Sample Read a File

Current Data File: 20090506114322

(0) SPP_200 Setup Debug

00 d 11:45:20 (42321sec)

11:43:23 12:30:00 12:57:50

Pump Relay Control This light indicates that Conc/Vol Conc has turned the relay on.

ON OFF PADS thinks it has

Relay On There is no direct feed-back from the probe.
SPP_200 Data

Manual Slow Medium Fast
11:43:23 12:30:00 12:57:50

Selectable Charts

SPP_200 # Conc (#/cm^3)

Enable

No Fault

v2.7.2

Enabled

COM Port 5

Auto-Scale

X-Axis

Y-Axis

Particle Size

Log-Scale

Normalized

Auto-Scale

0.34

0.36

0.32

0.34

0.36

0.38

0.40

0.42

0.44

0.46

0.48

0.50

0.52

0.54

0.56

0.58

0.60

0.62

0.64

0.66

0.68

0.70

0.72

0.74

0.76

0.78

0.80

0.82

0.84

0.86

0.88

0.90

0.92

0.94

0.96

0.98

1.00

1.02

1.04

1.06

1.08

1.10

1.12

1.14

1.16

1.18

1.20

1.22

1.24

1.26

1.28

1.30

1.32

1.34

1.36

1.38

1.40

1.42

1.44

1.46

1.48

1.50

1.52

1.54

1.56

1.58

1.60

1.62

1.64

1.66

1.68

1.70

1.72

1.74

1.76

1.78

1.80

1.82

1.84

1.86

1.88

1.90

1.92

1.94

1.96

1.98

2.00

2.02

2.04

2.06

2.08

2.10

2.12

2.14

2.16

2.18

2.20

2.22

2.24

2.26

2.28

2.30

2.32

2.34

2.36

2.38

2.40

2.42

2.44

2.46

2.48

2.50

2.52

2.54

2.56

2.58

2.60

2.62

2.64

2.66

2.68

2.70

2.72

2.74

2.76

2.78

2.80

2.82

2.84

2.86

2.88

2.90

2.92

2.94

2.96

2.98

3.00

3.02

3.04

3.06

3.08

3.10

3.12

3.14

3.16

3.18

3.20

3.22

3.24

3.26

3.28

3.30

3.32

3.34

3.36

3.38

3.40

3.42

3.44

3.46

3.48

3.50

3.52

3.54

3.56

3.58

3.60

3.62

3.64

3.66

3.68

3.70

3.72

3.74

3.76

3.78

3.80

3.82

3.84

3.86

3.88

3.90

3.92

3.94

3.96

3.98

4.00

4.02

4.04

4.06

4.08

4.10

4.12

4.14

4.16

4.18

4.20

4.22

4.24

4.26

4.28

4.30

4.32

4.34

4.36

4.38

4.40

4.42

4.44

4.46

4.48

4.50

4.52

4.54

4.56

4.58

4.60

4.62

4.64

4.66

4.68

4.70

4.72

4.74

4.76

4.78

4.80

4.82

4.84

4.86

4.88

4.90

4.92

4.94

4.96

4.98

5.00

5.02

5.04

5.06

5.08

5.10

5.12

5.14

5.16

5.18

5.20

5.22

5.24

5.26

5.28

5.30

5.32

5.34

5.36

5.38

5.40

5.42

5.44

5.46

5.48

5.50

5.52

5.54

5.56

5.58

5.60

5.62

5.64

5.66

5.68

</div

DLL Particle Analysis and Display System

Program Configure

Current Data File: 20090506114322

① SPP_200 Setup Debug

Pump Relay Control This light indicates that Conc/Vol Conc

OFF PADS thinks it has

turned the relay on.
SPP_200 Data There is no direct feed-

back from the probe.

Number Conc Sample Flow

(# / cm³) (vol cm³/s)

611.27 1.26

Volume Conc Sheath Flow
(cm³ / cm³)(vol cm³/s)

5.53 14.23

MVD (μm) Laser Ref (V)

0.29 9.04

ED (μm) Electronics

Temp (C)

0.3 40

Aux Analog 1 Avg Transit

-0.01 276

Hi Gain

Baseline (V) ADC Overflow

0.04 0

Mid Gain

Baseline (V) Pump Flow

0.34 0 OK

Lo Gain

Baseline (V) .11

0.34 .12 .13 .14

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

.15 .16 .17 .18

0cc


00 d 11:44:37 (42278sec)

Manual Slow Medium Fast

11:43:23 12:30:00 12:57:50

Selectable Charts
 No Fault v2.7.2


SPP_200 Volume Conc (μm³/cm³)




Enable

Enabled COM Port 5

SPP_200 # Conc (# /cm³)



X-Axis Auto-Scale



Particle Size

Y-Axis Normalized

Log-Scale

Auto-Scale


DLL Particle Analysis and Display System

Program Configure

Read a File

Current Data File: 20090506114322


(0) SPP_200 Setup Debug





Pump Relay Control: This light indicates that PADS thinks it has turned the relay on.

Relay On ● There is no direct feed-back from the probe.

SPP_200 # Conc/Vol Conc (#/cm^3)



SPP_200 Volume Conc (μm³/cm³)



Y-Axis Log-Scale Normalized
Auto-Scale

Particle Size

SDT Particle Analysis and Display System

Program Configure

Current Data File: 20050506114322

(0) SPP_200 Setup Debug

Pump Relay Control This light indicates that PADS thinks it has turned the relay on.

ON OFF Relay On There is no direct feed-back from the probe.

SPP_200 Data

Number Conc

(# / cm³)

477.9 Volume Conc

(um³ / cm³)

1.26 Sheath Flow

(vol cm³/s)

1.26

191.82 MVD (um)

14.34 Laser Ref (V)

1.75 ED (um)

9.04 Electronics

Temp (C)

39.1 Aux Analog 1

Avg Transit

-0.01 Hi Gain

237 Baseline (V)

0 ADC Overflow

0.06 Mid Gain

0 Pump Flow

OK

0.37 Lo Gain

40 Baseline (V)

11 Pump Flow

OK

0.35 Lo Gain

40 Baseline (V)

11 Pump Flow

OK

0.35 Lo Gain

40 Baseline (V)

11 Pump Flow

OK

0cc
00 d 11:44:02 (42242sec)
12:57:50


Manual Slow Medium Fast 11:43:23 12:30:00 12:57:50

Enabled COM Part 5 No Fault v2.7.2


Selectable Charts

Enable


SPP_200 # Conc (#/cm³)



SPP_200 volume Conc (um³/cm³)



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




X-Axis Auto-Scale



X-Axis Auto-Scale



X-Axis Auto-Scale




DAI Particle Analysis and Display System



Program Configure

Current Data File: 20090506114322

(0) SPP_200 Setup Debug



v2.7.2



DMT Particle Analysis and Display System

Program Configure

Sample **Read a File**

Current Data File: 20090506114322

(0) SPP_200 Setup Debug

Pump Relay Control This light indicates that Conc/Vol Conc

ON OFF PADS thinks it has turned the relay on.

SPP_200 Data back from the probe.

Number Conc (# / cm ³)	Sample Flow (vol cm ³ /s)	Volume Conc (cm ³ / cm ³)(vol cm ³ /s)	MVD (μm)	ED (μm)	Aux Analog 1	Avg Transit
2064.38	1.26	2064.38 μm ³ / cm ³ (vol cm ³ /s)	5.53	0.22	-0.01	194
	Sheath Flow (#/cm ³)	Laser Ref (V)	9.96	8.96	Hi Gain Baseline (V)	ADC Overflow
	#/cm ³	Electronics	14.05	14.05	0.05	0
		Temp (C)	37.1	37.1	0.34	0.34
					Mid Gain Baseline (V)	Pump Flow OK
					Lo Gain Baseline (V)	

SPP_200 # Conc (#/cm³)

SPP_200 # Conc (#/cm³)

Selectable Charts

Enable **Enabled** **No Fault** v2.7.2

COM Port 5

X-Axis Auto-Scale

Particle Count

Y-Axis Log-Scale Normalized

Bins

Particle Size

