


Certificate of Calibration

Report Number: 137.36278**PO Number** UND01-0000005706**SensorModel:** A3**Sales Order Number:** A084124

The following designated Dew Point Hygrometer was calibrated on the below date against EdgeTech Model DewMaster Standard Serial Number 3013CRX and meets the published accuracy specifications. The DewMaster was calibrated against a Model DewMaster Primary Standard Hygrometer Serial Number 3H2579MCR. The Primary Standard Hygrometer was calibrated directly by NIST on 22-JAN-08, NIST Test Report Number 276089-08. The primary standard is next due for calibration on 22-JAN-10. The procedure used for calibration is A31922 Rev. A. This calibration was performed in full compliance with ANSI/NCSS Z540-1-1994 requirements.

The Model 137**Accuracy is** +/- 0.50 Deg. C**Customer ID Number :****Dew Point Calibration is in :** Celsius**As Found Standard :** NEW**As Found S/N 1776 :** NEW**As Left Standard:** 14.49**As Left S/N 1776 :** 14.47**Date Calibrated** 5/15/2008**Calibrated By :**
Randy Taylor
Service and Repair Manager**Important Notes :** Calibrated with a 15 ft. remote sensor cable.